

Certificate for the Proficiency Test of Olfactometry 2019

This certificate was issued to confirm the participation of the laboratory:

Odometric
6700 Weyler
Belgique

with the coding number: 121


in the international proficiency test for olfactometry, held in June 2019. The aim of the test was to determine the precision r and accuracy A_{od} according to EN 13725:2003. Results of the participant are as follows:

Quality Parameter assessed using 1-Butanol in Nitrogen	Results (requirements according to EN 13725)
$A_{od} = 0,031$	complies with requirement: $A \leq 0,217$
$r = 0,122$	complies with requirement: $r \leq 0,477$

The odorant provided was: 1-butanol in nitrogen.

Number of participating laboratories: 41

Kiel, September 4th, 2019


(Björn Maxeiner - Director, Olfasense GmbH)

